







Test Report

Report No.: QDHL2112514553MD

Applicant: HAINAN FOSUN MEDICAL TRADE LIMITED

COMPANY

Manufacturer:

Sample Description: DISPOSABLE NITRILE EXAM GLOVES

Sample Type: M

Test Type: SUBMITTED BY CLIENT

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.



SGS Center, No.143, Zhuzhou Road, Laoshan District, Qingdao, Shandong, China 266101 t (86-532) 68999888 www.sgsgroup.com.cn sgs.china@sgs.com







Statement

- 1. The report is considered invalidated if no testing seal of SGS.
- 2. The test report cannot be reproduced in any way, except in full content, without prior approval in writing by the laboratory.
- 3. The copy report is considered invalidated if no red testing seal of SGS.
- 4. The report is considered invalidated if no approval signature.
- 5. The report is considered invalidated if altered.
- 6. Should you have any queries or objection to the test report, please contact us in writing within 15 days after receiving the report.
- 7. The results shown in this test report refer only to the sample(s) tested.
- 8. The sample information and sample(s) was/were submitted and certified by the client, SGS quoted the information with no responsibility as to the accuracy, adequacy and/or completeness.

Company Name:

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Company Address:

SGS Center, No.143, Zhuzhou Road, Laoshan District, Qingdao, Shandong, China.

Tel: +86(0)532-68999187

Fax: +86(0)532-80991952

Zip: 266101











Report No.: QDHL2112514553MD Page 1 of 4

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Test Report Home Page

Sample Description	DISPOSABLE NITRILE EXAM GLOVES		QDHL2112514553MD			
	Submitted by Random Applicant (√) sampling (——)	- Sample No.				
Brand	1	Type/Specifications	M			
Applicant	HAINAN FOSUN MEDICAL TRADE LIMITED COMPANY	Test Type	SUBMITTED BY CLIENT			
Applicant Address	ROOM 203-1,205-10,207-2, INTERNATIONAL BUSINESS CENTER, CROSS-BORDER E- COMMERCE INDUSTRIAL PARK, HAIKOU COMPREHENSIVE BONDED ZONE, HAIKOU, HAINAN	Lot No.	20211031			
Manufacturer	1	Inspection Report No.				
Inspected Company		Manufacture Date	2021/10/31			
Sampled By		Sample Quantity	220PCS 150000PCS QINGDAO			
Sampled Location		Lot Quantity				
Sampled Date		Testing Location				
Sample Received Date	DEC.03,2021	Testing Period	DEC.03,2021 TO DEC.15,2021			
Test items	Water tightness test for detection of holes					
Testing Accordance	EN 455-1:2020 Medical Gloves for Single Use – Part 1: Requirements and Testing for Freedom from Holes Clause 5.1					
Test conclusion	This report only provides the test results and individual assessment, details please refer to the following page(s).					
			Issue date: DEC.15,2021			









中国认可 国际互认 **TESTING CNAS L0604**

Report No.: QDHL2112514553MD Page 2 of 4

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Remark	 "——"=Not Applicable, "/"=Blank. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested. This document cannot be used for publicity, without prior written approval of the SGS.
--------	---

Approver/Date:













Report No.: QDHL2112514553MD Page 3 of 4

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Test Results

N	10.	Test Items	Test Method	Requirement	Test Result	Assessment	Remark
	1	Water tightness test for detection of holes	EN 455-1: 2020 Clause 5.1	Sample quantity: 200pcs AQL: 1.5 Ac: 7 Re: 8	Found: 2pcs	Pass	/









中国认可 国际互认 **TESTING CNAS L0604**

Report No.: QDHL2112514553MD Page 4 of 4

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

Test Report Picture Page

Picture and Description



Sample Description

Color: Blue; Size: M; Material:Nitrile; Textured fingertip; Examination Gloves; Powder-Free Gloves

Type/Specifications/Others

End of Report



SGS Center, No.143, Zhuzhou Road, Laoshan District, Qingdao, Shandong, China 266101 t (86-532) 68999888 www.sgsgroup.com.cn sgs.china@sgs.com